

**High Volume Air Sampler Calibration Worksheet**

**Project Title:** Design and Construction of Tsuen Wan Drainage Tunnel  
**Monitoring Location:** Ho Fung College  
**Calibration Date:** 08-Apr-09  
**Calibration Due Date:** 08-Jun-09  
**Time:** 17:15

Sampler Model:	BM2000HX
Serial No.:	4994
Calibrator Orifice no.:	517N
Slope (m):	2.02953
Intercept (b):	-0.01939
Correction coeff. (r)	0.9999


$$\text{Flow (corrected)} = \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}}$$

Standard pressure (mmHg) Pstd:	749.3
Standard temp. (K) Tstd:	296.0
Calibration pressure (mmHg) Pa:	762.4
Calibration temp. (K) Ta:	294.3

$$Q_{std} = \frac{1}{m} \times \left( \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}} - b \right)$$

Sample no.	Pressure Drop (H), inch	Flow (corrected), m <sup>3</sup> /min	Actual flow rate (Qstd), m <sup>3</sup> /min	Flow indication (I), arbitrary
1	12.3	3.560	1.764	57.0
2	10.2	3.242	1.607	52.0
3	7.6	2.798	1.388	41.0
4	4.7	2.201	1.094	28.0
5	3.4	1.872	0.932	22.0

Correlation Coefficient : 0.9987



Remark  
 1HPa = 0.750062 mmHg

**Calibrated by:** Mak Kei Ho  
 ( *MKH* )

**Date:** 16-4-09

**Checked by:** Tang Hiu Yeung  
 ( *THY* )

**Date:** 16-4-09

**High Volume Air Sampler Calibration Worksheet**

**Project Title:** Design and Construction of Tsuen Wan Drainage Tunnel  
**Monitoring Location:** Heng Hoi Chi Hong Ship Temple  
**Calibration Date:** 08-Apr-09  
**Calibration Due Date:** 08-Jun-09  
**Time:** 16:40

Sampler Model:	BM2000HX
Serial No.:	5875
Calibrator Orifice no.:	517N
Slope (m):	2.02953
Intercept (b):	-0.01939
Correction coeff. (r)	0.9999


Standard pressure (mmHg) Pstd:	749.3
Standard temp. (K) Tstd:	296.0
Calibration pressure (mmHg) Pa:	762.4
Calibration temp. (K) Ta:	294.3

$$\text{Flow (corrected)} = \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}}$$

$$Q_{std} = \frac{1}{m} \times \left( \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}} - b \right)$$

Sample no.	Pressure Drop (H), inch	Flow (corrected), m <sup>3</sup> /min	Actual flow rate (Qstd), m <sup>3</sup> /min	Flow indication (I), arbitrary
1	12.2	3.545	1.756	55.0
2	9.5	3.129	1.551	46.0
3	7.5	2.780	1.379	39.0
4	4.5	2.153	1.070	28.0
5	3.0	1.758	0.876	22.0

Correlation Coefficient : 0.9980



Remark  
 1HPa = 0.750062 mmHg

**Calibrated by:** Mak Kei Ho  
 ( Ho )

**Date:** 16-4-09

**Checked by:** Tang Hiu Yeung  
 ( Yeung )

**Date:** 16-4-09

**High Volume Air Sampler Calibration Worksheet**

**Project Title:** Design and Construction of Tsuen Wan Drainage Tunnel  
**Monitoring Location:** Greenview Terrance  
**Calibration Date:** 08-Apr-09  
**Calibration Due Date:** 08-Jun-09  
**Time:** 12:20

Sampler Model:	TE5005X
Serial No.:	0646
Calibrator Orifice no.:	517N
Slope (m):	2.02953
Intercept (b):	-0.01939
Correction coeff. (r)	0.9999


Standard pressure (mmHg) Pstd:	749.3
Standard temp. (K) Tstd:	296.0
Calibration pressure (mmHg) Pa:	762.4
Calibration temp. (K) Ta:	294.3

$$\text{Flow (corrected)} = \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}}$$

$$Q_{std} = \frac{1}{m} \times \left( \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}} - b \right)$$

Sample no.	Pressure Drop (H), inch	Flow (corrected), m <sup>3</sup> /min	Actual flow rate (Qstd), m <sup>3</sup> /min	Flow indication (I), arbitrary
1	13.0	3.660	1.813	58.0
2	11.3	3.412	1.691	53.0
3	9.0	3.045	1.510	43.0
4	4.8	2.224	1.105	26.0
5	3.3	1.844	0.918	18.0

Correlation Coefficient : 0.9991



Remark  
 1HPa = 0.750062 mmHg

**Calibrated by:** Mak Kei Ho  
 ( *Ho* )

**Date:** 16-4-09

**Checked by:** Tang Hiu Yeung  
 ( *Hy* )

**Date:** 16-4-09

**High Volume Air Sampler Calibration Worksheet**

**Project Title:** Design and Construction of Tsuen Wan Drainage Tunnel  
**Monitoring Location:** Long Beach Gardan  
**Calibration Date:** 08-Apr-09  
**Calibration Due Date:** 08-Jun-09  
**Time:** 11:55

Sampler Model:	TE5005X
Serial No.:	0390
Calibrator Orifice no.:	517N
Slope (m):	2.02953
Intercept (b):	-0.01939
Correction coeff. (r)	0.9999


Standard pressure (mmHg) Pstd:	749.3
Standard temp. (K) Tstd:	296.0
Calibration pressure (mmHg) Pa:	762.4
Calibration temp. (K) Ta:	294.3

$$\text{Flow (corrected)} = \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}}$$

$$Q_{std} = \frac{1}{m} \times \left( \sqrt{H \times \frac{Pa}{Pstd} \times \frac{Tstd}{Ta}} - b \right)$$

Sample no.	Pressure Drop (H), inch	Flow (corrected), m <sup>3</sup> /min	Actual flow rate (Qstd), m <sup>3</sup> /min	Flow indication (I), arbitrary
1	11.9	3.501	1.735	54.0
2	10.5	3.289	1.630	50.0
3	8.0	2.871	1.424	41.0
4	4.8	2.224	1.105	30.0
5	3.5	1.899	0.945	22.0

Correlation Coefficient : 0.9987



Remark  
 1HPa = 0.750062 mmHg

**Calibrated by:** Mak Kei Ho  
 ( *Mak Kei Ho* )

**Date:** 16-4-09

**Checked by:** Tang Hiu Yeung  
 ( *Tang Hiu Yeung* )

**Date:** 16-4-09



# Calibration Certificate

Certificate No. **83174**

Page 1 of 4 Pages

**Customer :** Hyder Consulting Limited

**Address :** 47/F., Hopewell Centre, 183 Queens Road East, Wanchai, Hong Kong

**Order No. :** Q81258

**Date of receipt :** 9-Jul-08

## Item Tested

**Description :** Sound Level Meter

**Manufacturer :** B&K

**Model :** 2238

**Serial No. :** 2448529

## Test Conditions

**Date of Test :** 9-Jul-08

**Supply Voltage :** --

**Ambient Temperature :** (23 ± 3)°C

**Relative Humidity :** (50 ± 25) %

## Test Specifications

Calibration check.

Calibration procedure : Z01.

## Test Results

All results were within the IEC 651 Type 1, IEC 804 Type 1 & IEC 1260 Class 1 specification.

The results are shown in the attached page(s).

Main Test equipment used:

<u>Equipment No.</u>	<u>Description</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceable to</u>
S017	Multi-Function Generator	C081456	18-Mar-09	SCL-HKSAR
S024	Sound Level Calibrator	71791	16-Jul-08	NIM-PRC & SCL-HKSAR

The values given in this Calibration Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environmental changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Hong Kong Calibration Ltd. shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration are traceable to International System of Units (SI).

The test results apply to the above Unit-Under-Test only

**Calibrated by :**   
P.F. Wong

**Approved by :**   
Alan Chu

**Date:** 10-Jul-08

This Certificate is issued by:

Hong Kong Calibration Ltd.

Unit 8B, 24/F., Well Fung Industrial Centre, No. 58-76, Ta Chuen Ping Street, Kwai Chung, NT, Hong Kong.

Tel: 2425 8801 Fax: 2425 8646



# Calibration Certificate

Certificate No. **83174**

Page 2 of 4 Pages

Results :

## 1. SPL Accuracy

UUT Setting				Applied Value (dB)	UUT Reading (dB)
Range	Freq. Wgt.	Bandwidth	Center Freq.		
20 ~ 100	A	BB/F	--	94.03	93.9
	A	BB/S	--		93.9
	C	BB/F	--		93.9
40 ~ 120	A	BB/F	--	94.03	94.0
	A	BB/F	--	113.97	113.8
40 ~ 120	--	1/3 - Oct./F	1 kHz	94.03	94.0
				113.97	113.8
40 ~ 120	--	1/1 - Oct./F	1 kHz	94.03	94.0
				113.97	113.8

IEC 651 Type 1 Spec. :  $\pm 0.7$  dB

Uncertainty :  $\pm 0.1$  dB

## 2. Level Stability : 0.0 dB

IEC 651 Type 1 Spec. :  $\pm 0.3$  dB

Uncertainty :  $\pm 0.01$  dB

## 3. Linearity

### 3.1 Level Linearity

UUT Range (dB)	Applied Value (dB)	UUT Reading (dB)	Variation (dB)	IEC 651 Type 1 Spec. (Primary Indicator Range) $\pm 0.7$ dB
130	114.0	114.2	0.2	
130	104.0	104.2	0.2	
120	94.0	94.0 (Ref.)	--	
110	84.0	83.8	0.2	
100	74.0	73.9	0.1	
90	64.0	63.9	0.1	
80	54.0	54.0	0.0	

Uncertainty :  $\pm 0.1$  dB



# Calibration Certificate

Certificate No. **83174**

Page 3 of 4 Pages

## 3.2 Differential level linearity

UUT Range (dB)	Applied Value (dB)	UUT Reading (dB)	Variation (dB)	IEC 651 Type 1 Spec.
120	84.0	83.8	0.2	± 0.4 dB
	94.0	94.0 (Ref.)	--	
	95.0	95.0	0.0	± 0.2 dB
	104.0	104.2	0.2	± 0.3 dB
	105.0	105.2	0.2	± 1.0 dB

Uncertainty : ± 0.1 dB

## 4. Frequency Weighting

A weighting

Frequency	Attenuation (dB)	IEC 651 Type 1 Spec.
31.5 Hz	- 39.5	- 39.4 dB, ± 1.5 dB
63 Hz	- 26.4	- 26.2 dB, ± 1.5 dB
125 Hz	- 16.5	- 16.1 dB, ± 1 dB
250 Hz	- 9.0	- 8.6 dB, ± 1 dB
500 Hz	- 3.5	- 3.2 dB, ± 1 dB
1 kHz	0.0 (Ref)	0 dB, ± 1 dB
2 kHz	+ 1.4	+ 1.2 dB, ± 1 dB
4 kHz	+ 1.2	+ 1.0 dB, ± 1 dB
8 kHz	- 0.8	- 1.1 dB, + 1.5 dB ~ -3 dB
16 kHz	- 6.3	- 6.6 dB, + 3 dB ~ - ∞

Uncertainty : ± 0.1 dB

## 5. Time Averaging

Applied Burst duty Factor	Applied Leq Value (dB)	UUT Reading (dB)	IEC 804 Type 1 Spec.
continuous	40.0	40.0	--
1/10	40.0	40.0	± 0.5 dB
1/10 <sup>2</sup>	40.0	40.0	
1/10 <sup>3</sup>	40.0	40.0	
1/10 <sup>4</sup>	40.0	39.5	± 1.0 dB

Uncertainty : ± 0.1 dB



# Calibration Certificate

Certificate No. **83174**

Page 4 of 4 Pages

## 6. Filter Characteristics

### 6.1 1/1 – Octave Filter

Frequency	Attenuation (dB)	IEC 1260 Class 1 Spec. (dB)
125 Hz	- 64.2	< - 61
250 Hz	- 45.0	< - 42
500 Hz	- 21.1	< - 17.5
707 Hz	- 3.8	- 2 ~ - 5
1 kHz (Ref)	--	--
1.414 kHz	- 3.7	- 2 ~ - 5
2 kHz	- 20.8	< - 17.5
4 kHz	- 44.6	< - 42
8 kHz	- 63.8	< - 61

Uncertainty :  $\pm 0.25$  dB

### 6.2 1/3 – Octave Filter

Frequency	Attenuation (dB)	IEC 1260 Class 1 Spec.(dB)
326 Hz	- 64.7	< - 61
530 Hz	- 47.3	< - 42
772 Hz	- 22.5	< - 17.5
891 Hz	- 3.6	+ 0.3 ~ - 5.0
1 kHz (Ref)	--	--
1.122 kHz	- 3.5	+ 0.3 ~ - 5.0
1.296 kHz	- 22.4	< - 17.5
1.887 kHz	- 46.9	< - 42
3.070 kHz	- 65.2	< - 61

Uncertainty :  $\pm 0.25$  dB

- Remarks :
1. UUT : Unit-Under-Test
  2. The uncertainty claimed is for a confidence probability of not less than 95%.
  3. Atmospheric pressure : 1 000 hPa.

----- END -----



# Calibration Certificate

Certificate No. **83175**

Page 1 of 4 Pages

**Customer :** Hyder Consulting Limited

**Address :** 47/F., Hopewell Centre, 183 Queens Road East, Wanchai, Hong Kong

**Order No. :** Q81258

**Date of receipt :** 9-Jul-08

## Item Tested

**Description :** Digital Sound Level Meter

**Manufacturer :** B&K

**Model :** Type 2236

**Serial No. :** 1774423

## Test Conditions

**Date of Test :** 9-Jul-08

**Supply Voltage :** --

**Ambient Temperature :** (23 ± 3)°C

**Relative Humidity :** (50 ± 25) %

## Test Specifications

Calibration check.

Calibration procedure : Z01.

## Test Results

All results were within the IEC 651 Type 1, IEC 804 Type 1 & IEC 1260 Class 1 specification.

The results are shown in the attached page(s).

Main Test equipment used:


<u>Equipment No.</u>	<u>Description</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceable to</u>
S017	Multi-Function Generator	C081456	18-Mar-09	SCL-HKSAR
S024	Sound Level Calibrator	71791	16-Jul-08	NIM-PRC & SCL-HKSAR
S031	6½ dgt. Multimeter	76189	28-Dec-08	NIM-PRC

The values given in this Calibration Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environmental changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Hong Kong Calibration Ltd. shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration are traceable to International System of Units (SI).

The test results apply to the above Unit-Under-Test only

**Calibrated by :**   
P.F. Wong

**Approved by :**   
Alan Chu

**Date:** 10-Jul-08

This Certificate is issued by:

Hong Kong Calibration Ltd.

Unit 8B, 24/F., Well Fung Industrial Centre, No. 58-76, Ta Chuen Ping Street, Kwai Chung, NT, Hong Kong.

Tel: 2425 8801 Fax: 2425 8646

The copyright of this certificate is owned by Hong Kong Calibration Ltd.. It may not be reproduced except in full.

# Calibration Certificate

Certificate No. 83175

Page 2 of 4 Pages

Results :

## 1. SPL Accuracy

UUT Setting				Applied Value (dB)	UUT Reading (dB)
Range	Parameter	Frequency Wt.	Freq. Response		
20 - 100	SPL	dBA	F	94.03	94.0
			S		94.0
		dBC	F		94.1
		dBL	F		94.1
		1 kHz	F		94.0
40 - 120	SPL	dBA	F	94.03	94.0
		1 kHz	F		94.0
	SPL	dBA	F	113.97	114.0
			S		114.0
		dBC	F		114.0
		dBL	F		114.0
		1 kHz	F		114.0

IEC 651 Type 1 Spec. :  $\pm 0.7$  dB

Uncertainty :  $\pm 0.1$  dB

## 2. Level Stability : 0.0 dB

IEC 651 Type 1 Spec. :  $\pm 0.3$  dB

Uncertainty :  $\pm 0.01$  dB

## 3. Linearity

### 3.1 Level Linearity

UUT Range (dB)	Applied Value (dB)	UUT Reading (dB)	Variation (dB)	IEC 651 Type 1 Spec. (Primary Indicator Range)
130	114.0	114.1	0.1	$\pm 0.7$ dB
130	104.0	104.2	0.2	
120	94.0	94.0 (Ref.)	--	
110	84.0	83.9	0.1	
100	74.0	73.8	0.2	
100	64.0	63.8	0.2	
100	54.0	53.9	0.1	

Uncertainty :  $\pm 0.1$  dB



# Calibration Certificate

Certificate No. 83175

Page 4 of 4 Pages

## 6. Filter Response

Filter Setting	Attenuation (dB)	IEC 1260 Class 1 Spec.
125 Hz	- 64.0	< - 61
250 Hz	- 44.8	< - 42
500 Hz	- 20.8	< - 17.5
707 Hz	- 3.5	- 2 ~ - 5
1 kHz (Ref.)	0.0 (Ref.)	--
1.414 kHz	- 4.0	- 2 ~ - 5
2 kHz	- 21.2	< - 17.5
4 kHz	- 45.0	< - 42
8 kHz	- 64.3	< - 61

Uncertainty :  $\pm 0.2$  dB

- Remark : 1. UUT : Unit-Under-Test  
2. The uncertainty claimed is for a confidence probability of not less than 95%.  
3. Atmospheric Pressure : 1 000 hPa

----- END -----

# Calibration Certificate

Certificate No. **83175**

Page 3 of 4 Pages

## 3.2 Differential level linearity

UUT Range (dB)	Applied Value (dB)	UUT Reading (dB)	Variation (dB)	IEC 651 Type 1 Spec.
120	84.0	83.8	0.2	± 0.4 dB
	94.0	94.0 (Ref.)	--	
	95.0	95.1	0.1	± 0.2 dB
	104.0	104.2	0.2	± 0.3 dB
	105.0	105.0	0.0	± 1.0 dB

Uncertainty : ± 0.1 dB

## 4. Frequency Weighting

A weighting

Frequency	Attenuation (dB)	IEC 651 Type 1 Spec.
31.5 Hz	- 39.9	- 39.4 dB, ± 1.5 dB
63 Hz	- 26.6	- 26.2 dB, ± 1.5 dB
125 Hz	- 16.5	- 16.1 dB, ± 1 dB
250 Hz	- 8.9	- 8.6 dB, ± 1 dB
500 Hz	- 3.4	- 3.2 dB, ± 1 dB
1 kHz	0.0 (Ref)	0 dB, ± 1 dB
2 kHz	+ 1.3	+ 1.2 dB, ± 1 dB
4 kHz	+ 1.1	+ 1.0 dB, ± 1 dB
8 kHz	- 1.1	- 1.1 dB, + 1.5 dB ~ -3 dB
16 kHz	- 7.2	- 6.6 dB, + 3 dB ~ -∞

Uncertainty : ± 0.1 dB

## 5. Time Averaging

Applied Burst duty Factor	Applied Leq Value (dB)	UUT Reading (dB)	IEC 804 Type 1 Spec.
continuous	40.0	40.0	--
1/10	40.0	40.0	± 0.5 dB
1/10 <sup>2</sup>	40.0	39.9	
1/10 <sup>3</sup>	40.0	39.8	± 1.0 dB
1/10 <sup>4</sup>	40.0	39.2	

Uncertainty : ± 0.1 dB



輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No. : C090563

## *Certificate of Calibration*

*This is to certify that the equipment*

*Description : Acoustical Calibrator*

*Manufacturer : Bruel & Kjaer*

*Model No. : 4231*

*Serial No. : 1770806*

*has been calibrated for the specific items and ranges.  
The results are shown in the Calibration Report No. C090563.*

*The equipment is supplied by*

*Co. Name : Hyder Consulting Limited*

*Address : 47/F., Hopewell Centre, 183 Queen's Road East,  
Wanchai, Hong Kong*

*Date of Issue : 6 February 2009*

Certified by :

  
C F Leung

The test equipment used for calibration are traceable to the National Standards as specified in this report.  
This report shall not be reproduced except in full and with prior written approval from this laboratory.

Calibration and Testing Laboratory of Sun Creation Engineering Limited

c/o 4/F, Tsing Shan Wan Exchange Building, 1 Hing On Lane, Tuen Mun, New Territories, Hong Kong

Tel: 2927 2606

Fax: 2744 8986

E-mail: callab@suncreation.com

Website: www.suncreation.com



1412 Honour Ind. Centre  
6 Sun Yip St. Chai Wan  
Hong Kong

<b>CERTIFICATE OF CALIBRATION</b>	
<b>IN - HOUSE</b>	
Date Of Issue :	Serial No : IC 42b / /EL

Item Being Calibrated : **Turbidity Standards (Gelex)**      Date Of Calibration : 8/4/09  
 Item Stock No : Std1,2,3,4      Operator : K.K  
 Environment Temp. °C 22      Procedure No Used : IC 42 (Revision No. 0)  
 Primary Standards use 20, 100 and 800 NTU Formazin standards prepared fr 03681  
 Ref. Equip.used/ Stock No : Serial No. 215619

\*\*\*\*\*

Gelex Standards	Turbidity of standard solution used (NTU)	Measured Value (NTU)	R <sup>2</sup>	Requirement R <sup>2</sup>
0 - 10 NTU	1	1.05	1	> 0.996
	5	5.23		
	10	10.47		
10 - 100 NTU	20	20.1	0.9967	> 0.996
	50	52.8		
	80	79.9		
100 - 1000 NTU	100	99.7	0.993	> 0.996
	400	452		
	800	807		

**Comments :**      *The equipment and Gelex Standards complies / ~~does not comply~~ with the Manufacturer's recommendation.*

Input data checked by :       Certified by:   
 Operations Manager



1412 Honour Ind. Centre  
6 Sun Yip St, Chai Wan  
Hong Kong

**CERTIFICATE OF CALIBRATION**  
**IN - HOUSE**

Date Of Issue : \_\_\_\_\_ Serial No : IC 42a / / EL

21977/1

Item Being Calibrated : Turbidity Standards (Gelex) Date Of Calibration : 8/4/09  
 Item Stock No : Std1,2,3,4 Operator : K.K  
 Environment Temp. °C 22 Procedure No Used : IC 42 (Revision No. 0)  
 Primary Standards used 20, 100 and 800 NTU Formazin standards prepared fr 03681  
 Ref. Equip.used/ Stock No : Serial No. 215619

\*\*\*\*\*

Gelex Standards	Last assigned value Date: (NTU)	New measured value (NTU)	Agreement %	Requirement %
0 - 10 NTU	0	0	0	± 5
10 - 100 NTU	16.86	16.51	-2.08	± 5
100 - 1000 NTU	100	95.7	-4.30	± 5
100 - 1000 NTU	861	868	0.81	± 5

**Comments :** *The equipment and Gelex Standards complies / does not comply with the Manufacturer's recommendation.*

Input data checked by : \_\_\_\_\_

Certified by: \_\_\_\_\_  
Operations Manager

# CERTIFICATE OF ANALYSIS




Batch: HK0906207  
Date of Issue: 06/04/2009  
Client: HYDER CONSULTING LTD  
Client Reference:

## Calibration of DO System

Item : Multi-parameter Instrument / Mehrparameter-Meßgerät  
Model No. : WTW pH / Oxi 340i  
Serial No. : 08101283  
Equipment No.: --  
Calibration Method : This meter was calibrated in accordance with standard method APHA (18th Ed.) 4500-0C & G  
Date of Calibration : 06 April, 2009

### Testing Results :

Expected Reading	Recording Reading
4.70 mg/L	4.87 mg/L
6.64 mg/L	6.70 mg/L
8.52 mg/L	8.68 mg/L
Allowing Deviation	±0.2 mg/L

  
Ms Wong Wai Man, Alice  
Laboratory Manager - Hong Kong

# CERTIFICATE OF ANALYSIS



Batch: HK0906207  
Date of Issue: 06/04/2009  
Client: HYDER CONSULTING LTD  
Client Reference:

## Calibration of pH System

Item : Multi-parameter Instrument / Mehrparameter-MeBgerat  
Model No. : WTW pH / Oxi 340i  
Serial No. : 08101283  
Equipment No. : --  
Calibration Method : This meter was calibrated in accordance with standard method APHA (19th Ed.) 4500-H<sup>+</sup>B  
Date of Calibration : 06 April, 2009

### Testing Results :

Expected Reading	Recording Reading
4.00	4.11
7.00	7.02
10.0	9.85
Allowing Deviation	± 0.2

  
Ms Wong Wai Man, Alice  
Laboratory Manager - Hong Kong

# CERTIFICATE OF ANALYSIS




**Batch:** HK0906207  
**Date of Issue:** 06/04/2009  
**Client:** HYDER CONSULTING LTD  
**Client Reference:**

## Calibration of Thermometer

**Item :** Multi-parameter Instrument / Mehrparameter-MeBgerat  
**Model No. :** WTW pH / Oxi 340i  
**Serial No. :** 08101283  
**Equipment No. :** --  
**Calibration Method :** In-house Method  
**Date of Calibration :** 06 April, 2009

## Testing Results :

Reference Temperature (°C)	Recorded Temperature (°C)
22.0 °C	22.3 °C
33.0 °C	33.5 °C
Allowing Deviation	±2.0°C

  
Ms Wong Wai Man, Alice  
Laboratory Manager - Hong Kong